



# Certificate of Analysis

## Purified Water, USP Sterile Grade

Product Number: WPW-USP-2XL

Lot Number: 33003214  
Date of Manufacture: 2021-03-30  
Expiration Date: 2024-03-30  
Storage Temperature: 2 - 30 °C

### BASIS FOR RELEASE

Test	Method	Specification	Result
Appearance	Visual	Clear and Colorless Liquid	Clear and Colorless Liquid
Conductivity (Packaged)	USP <645>	≤ 5.0 μS/cm	0.62 μS/cm
Oxidizable Substances	USP	The Pink Color does not Completely Disappear	Pass
Sterility	USP <71>	No Growth	No Growth
TOC (Before Packaging)	USP <643>	< 500 ppb	45.50 ppb

This lot has been produced using Current Good Manufacturing Practices (cGMP). The solution meets all requirements for residual solvents USP <467>.

This product is not of animal origin and does not contain, make use of or involve animal components during its manufacturing process.

No antimicrobial or other substance has been added.

### STATEMENT FOR USE

This product is for further manufacturing, laboratory, or research use only.  
Not intended for direct human or veterinary parenteral administration or where prohibited by law.

### Approval

Quality Assurance – Print Name:	Quality Assurance – Sign Name:	Date:
Brenden Ok		29 Apr 2021

### Intermountain Life Sciences

8095 South Welby Park Drive | West Jordan, UT 84088 | USA  
Tel: 801-282-8610 | www.ilsgmp.com